



SAFETY AND BUILDINGS DIVISION  
Plumbing Product Review  
P.O. Box 2658  
Madison, Wisconsin 53701-2658  
**TTY: Contact Through Relay**  
**Jim Doyle, Governor**  
**Richard J. Leinenkugel, Secretary**

October 2, 2009

ARJO HUNTLEIGH  
GARY LOFGREN  
50 N GARY AVE  
ROSELLE IL 60172

ARJO HOSPITAL EQUIPMENT AB  
JOHN WENNBORG  
PO BOX 61  
S-241 21,ESLOV, SWEDEN

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: ARJO HOSPITAL EQUIPMENT AB  
Product Name: PARKER BATH  
Model Number(s): 410.00.01.05  
420.00.01.05  
Product File No: 20090361

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters Comm 82 through 84, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of February 2015.

This approval is contingent upon compliance with the following stipulation(s):

- The waste from this product must be discharged by means of an air-break or air-gap into a receptor through into a drain connected to the sanitary waste system.

This approval supersedes the approval issued on February 21, 2005 under product file number 20040535.

This approval letter shall be incorporated with your previously approved plans and/or specifications approved under product file number 20040535.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson  
Plumbing Product Reviewer  
Phone: 608-266-6742  
Fax: 608-267-9566  
E-mail: jerry.thompson@wi.gov